

```
#####  
# LMG calibration data file for sensors  
#  
# NOTE:  
# 1. In order for these calibrations to take affect, uwint and rv_tsg must  
# be restarted. (Remember, rv_tsg has parameters.)  
#  
# 2. Please enter serial numbers for all sensors  
#  
# 3. Remember, when you check this file back into RCS, use the  
# -u option. It MUST remain in /usr/local/packages/rvdas/config  
#  
# 4. The TSG calibration coefficients must be placed last in this file.  
#  
#  
#####  
# Ship - LMG or NBP  
SHIP LMG  
#  
#####  
#  
# Cruise ID (i.e. LMG0505)  
cruiseID LMG0802  
#  
#####  
# LM Gould radiometer calibrations  
# PSP ser#:31701F3 cal date: 25 Apr, 2007  
# PIR ser#:32031F3 cal date: 03 Apr. 2007  
# Instrument uVolts/W/m^2  
PSP 8.51  
PIR 3.91  
#  
#####  
# Instrument Vdark Calib_Factor (ser#:6394, cal date: 8/29/2006  
#instrument, Probe Dark(V), Calib Factor (Dry) (V/uE/cm^2sec)  
# Dark value measured in situ 28 Dec, 2007  
PAR -0.0063735 5.26  
#  
#####  
# Transmisometer (ser#: CST-891DR, cal date: 20Nov06)  
# Vdark Vref Path  
TRAN 0.057 4.829 0.25  
#  
#  
#####  
#
```

```

#
# LMG winches
#
# Scale conversion information for the science winches on the LMG.
# Sheave measurements made on 01/01/00.
# Wire Pull tests done on dates indicated
#
# Dush 4 winch    sheave diam=
# 9/16" wire    wire diam =
#     total circumference=
#             magnets    =
#     Payout Scale factor=
#     Tension Scale Factor=
#     operation limit= lb
#
# Dush xx winch    sheave diam= 28.125    .714m
# .680" wire    wire diam = 0.680    .017m
#     total circumference= 90.493" 2.297m
#             magnets    = 24
#     Payout Scale factor= 3.77    0.096m
#     Tension Scale Factor= 180
#     operation limit= 20,150 lb
#
#
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
# operation limit = d
#     a     b     c     d     e
LDU4  1     0.465 1  20718  0
LDU5  1     1     1  20150  0
LD11  1     1     1  5980   0
LWN1  1     1     1  5980   0
#SWNC -0.1  200   1.67 20718 -800
#PWNC  0.1  180   1.67 20150  0
#BWNC  0.1  62.5  1.67 5980   437.5
#WWNC -0.1  60    -1.67 5980   0
#
#
#####
#####
#### Note, TSG calibrations must be last in this file #####
#### Do not change the formating, only the values. Thanks #####
#####
##
#***** Calibration factors for SBE 21 S/N 3208 *****

```

\*\*\*\*\* Calibration Date of 30-Jun-05 2007 \*\*\*\*\*

# currently in use

# Temperature calibration factors

%TEMPERATURE%

g 0.00413343557

h 0.000615389618

i 0.0000197232448

j 0.00000135946639

fo 1000.000

\*

# conductivity calibration factors

%CONDUCTIVITY%

g -3.99168762

h 0.471887572

i -0.00055051605

j 0.0000513922485

p -0.0000000957

t 0.00000325

\*

\*\*\*\*\* Remote Temperature Probe SN # 4015 \*\*\*\*\*

\*\*\*\*\* Calibration Date of 11-May-07 \*\*\*\*\*

# external temperature calibration factors

%EXTERNAL TEMPERATURE%

g 0.0043665187

h 0.000627057431

i 0.0000215854061

j 0.00000173345987

fo 1000.000

\*

#

#

#