

```

#####
# LMG calibration data file for sensors
#
# NOTE:
# 1. In order for these calibrations to take affect, uwint and rv_tsg must
# be restarted. (Remember, rv_tsg has parameters.)
#
# 2. Please enter serial numbers for all sensors
#
# 3. Remember, when you check this file back into RCS, use the
# -u option. It MUST remain in /usr/local/packages/rvdas/config
#
# 4. The TSG calibration coefficients must be placed last in this file.
#
#####
# Ship - LMG or NBP
SHIP LMG
#
#####
# Cruise ID (i.e. LMG0810)
cruiseID LMG1006
#
#####
# LM Gould radiometer calibrations
# PSP ser#:31701F3 cal date: 20 Jan, 2010
# PIR ser#:32031F3 cal date: 20 Jan, 2010
# Instrument      uVolts/W/m^2
PSP      8.36
PIR      3.86
#
#####
# PAR Instrument Vdark Calib_Factor (ser#:6394, cal date: 10/27/2008
#instrument, Probe Dark(V), Calib Factor (Dry) (V/uE/cm^2sec)
PAR      .0004  5.8024
#
#####
# Transmisometer (ser#: CST-407DR, cal date: 13Oct09)
# Vdark Vref Path
TRAN     0.059  4.687  0.25
#
#
#####
# LMG winches
#
# Scale conversion information for the science winches on the LMG.
# Sheave measurements made on 01/01/00.
# Wire Pull tests done on dates indicated
#
# Dush 4 winch      sheave diam=
# 9/16" wire      wire diam  =
#      total circumference=

```

```

#           magnets      =
#       Payout Scale factor=
#       Tension Scale Factor=
#           operation limit=  lb
#
# Dush xx winch      sheave diam= 28.125   .714m
# .680" wire      wire diam  = 0.680     .017m
#       total circumference= 90.493"  2.297m
#           magnets      = 24
#       Payout Scale factor= 3.77     0.096m
#       Tension Scale Factor= 180
#           operation limit= 20,150 lb
#
#
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
# operation limit = d
#       a      b      c      d      e
LDU4   1      0.465  1  20718  0
LDU5   1      1      1  20150  0
LD11   1      1      1  5980   0
LWN1   1      1      1  5980   0
#SWNC  -0.1   200    1.67  20718  -800
#PWNC  0.1    180    1.67  20150  0
#BWNC  0.1    62.5   1.67  5980   437.5
#WWNC  -0.1   60     -1.67  5980   0
#
#
#####
#####
#### Note, TSG calibrations must be last in this file #####
#### Do not change the formating, only the values. Thanks #####
#####
##
***** Calibration factors for SBE 21 S/N 3208 *****
***** Calibration Date of 30-Jun-05 2007 *****
# currently in use
# Temperture calibration factors
#%TEMPERTURE%
#g 0.00413343557
#h 0.000615389618
#i 0.0000197232448
#j 0.00000135946639
#fo 1000.000
#*

# conductivity calibration factors
#%CONDUCTIVITY%
#g -3.99168762
#h 0.471887572
#i -0.00055051605
#j 0.0000513922485
#p -0.0000000957
#t 0.00000325

```

```
##
#***** Remote Temperature Probe SN # 4015 *****
#***** Calibration Date of 11-May-07 *****
# external temperature calibration factors
#%EXTERNAL TEMPERATURE%
#g 0.0043665187
#h 0.000627057431
#i 0.0000215854061
#j 0.00000173345987
#fo 1000.000
##
#
#
#
```