

```
#####  
#  
# NBP calibration data file for sensors  
#  
# NOTE:  
# 1. In order for these calibrations to take affect, uwint must  
# be restarted.  
#  
# 2. Please enter serial numbers for all sensors  
#  
# 3. Remember, when you check this file back into RCS, use the  
# -u option. It MUST remain in /usr/local/packages/rvdas/config  
#  
# Revised August 10, 2000 K. Gavahan  
# - Initial revision.  
#  
# Revised...  
#  
#####  
#  
# Ship - LMG or NBP  
#  
SHIP NBP  
#  
#####  
#  
# Science specific information  
#  
VESSEL: NBP  
TITLE: Stock  
NUMBER: NBP0607C  
START_DATE: 10/10/2006  
END_DATE: 10/24/2006  
CHIEF_SCIENTIST: Joann Stock  
PARTICIPATING SCIENTISTS:  
#  
#-----  
# Data specific information  
#  
# Base file name for data files  
BASE_FILE: NBP0607C  
#  
# NAVIGATION LOGGERS - loggers and data directory  
NAV_LOGGERS: l_gyr,l_pcode,l_seap  
DATA_LOC_NAV: /data/logger/nav  
#  
# UNDERWAY LOGGERS - loggers and data directory  
UW_LOGGERS: l_met,l_sim,l_tsg,l_bathy  
DATA_LOC_UWAY: /data/logger/uw  
#  
NETWORK: science  
#  
#  
LOGGER_LOC: /usr/local/packages/rvdas/bin  
#
```

```

#
#####
#
# Geophysical information
#
# The value for YEAR can be obtained by running /usr/local/bin/decimal_year.
# It should be updated everytime the gravity offset is updated.
#
# Gravity offset information
#
CRUISE_ID: NBP0607C
#GRAV_OFFSET: 972326.93 on 0401.
#GRAV_OFFSET: 972327.68 on 0402.
#GRAV_OFFSET: 972304.02 on 0501
#GRAV_OFFSET: 972305.39 on 0501b
#GRAV_OFFSET: 972308.37 on 0506
#GRAV_OFFSET: 972310.34 on 0601
#GRAV_OFFSET: 972311.94
#GRAV_OFFSET: 972309.88 on 0602
#GRAV_OFFSET: 972311.96
#GRAV_OFFSET: 972312.25 on 0603
#GRAV_OFFSET: 972311.95
#GRAV_OFFSET: 972318.07
#GRAV_OFFSET: 972319.71
GRAV_OFFSET: 972320.26
YEAR: 2006.77
#
#####
#
# Gravity QC
#
# LOCATION      : where the final data file is located
# GAP           : minimum allowable time gap      (in seconds)
# FIELD         : field where value can be found (starting at 1 after timestamp)
# NAME          : name of the field
# MIN           : minimum allowable value         (XXX means ignore)
# MAX           : maximum allowable value         (XXX means ignore)
# ROC           : maximum allowable rate of change (XXX means ignore)
# DELIMITER     : the delimiter for the body of data items (SPACE, COMMA, . . .)
#
#TAG LOCATION GAP FIELD NAME MIN MAX ROC DELIMITER
GRQC /data/current_cruise/geopdata/GRV/ 30 1 GRAVITY 4000 12000 100 SPACE
#####
#
# NBP PSP and PIR coefficients
# PSP serial number: 33090F3      Cal: 18 Jan 2006
# PIR serial number: 33023F3      Cal: 18 Jan 2006
#
# PSP and PIR assumed to be cal number * 10^-6
#
# Calculations make in the programs....
# PIR = mV / ( coeff V/wm^-2 * 10^3mV/V)
# PSP = mV / ( coeff v/wm^-2 * 10^3mV/V)
#
# For example, program will calculate

```

```

#      PIR coefficient = 1/ (3.92x10^-6 * 10^3) = 255.1
#      PSP coefficient = 1/(8.05x10^-6 * 10^3) = 124.22
#
PSP1 8.11
PIR1 3.86
#####
#
# NBP met
#
# PAR serial number: 6356
# PAR Calibration Factor = 6.27 V/uE/cm^2sec
# PAR Probe Dark offset = 0.1 mV
#
# old PAR      1/6.30 (Dry V/uE/cm^2sec) 0.0021 (Probe Dark in Volts)
#
# PAR      1/6.06 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)
# PAR      1/6.10 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)
# PAR      1/6.10 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)
# PAR      1/6.27 (Dry V/uE/cm^2sec) 0.0001 (Probe Dark in Volts)
#
PAR 0.15949 0.0001
#
#####
#
# Transmissiometer
# Serial number CST-557DR
# Date Calibrated 04/07/2005
#
# % transmission = (vsig - vd) / (vref - vd)
#
#   vd = 0.059
#   vref = 4.708
#
#   = (vsig - 0.058) / (4.772 - 0.058)
#
#   Vdark   Vref
TRAN 0.060  4.707
#
#####
#
# Engineering
#
# RPM pitch rudder
SENG 25.  10.  3400.  2500.  20.
PENG 25.  10.  3400.  2500.  20.
# Roll and Pitch Pot
POPI 4.0  4.0
# Seawater flow meter
# swfl *c1  +c2
SWFL 48.0  0
#
#####
#
# NBP winches
#

```

```

# Scale conversion information for the science
# winches on the NBP. Sheave measurements made
# on 01/01/00. Wire Pull tests done on dates
# indicated
#
# stbd winch   sheave diam= 28.125"  .714m
# 9/16" wire   wire diam  = 0.5625"  .014m
#             total circumference= 90.124" 2.289m
#             magnets      = 24
#             Payout Scale factor= 3.755   .095
#             Tension Scale Factor= 200
#             operation limit= 20,718 lb
#
# port winch   sheave diam= 28.125   .714m
# .680" wire   wire diam  = 0.680    .017m
#             total circumference= 90.493" 2.297m
#             magnets      = 24
#             Payout Scale factor= 3.77    0.096m
#             Tension Scale Factor= 180
#             operation limit= 20,150 lb
#
# baltic winch sheave diam= 12.125   .308m
# .322" wire   wire diam  = 0.322    .008m
#             total circumference= 39.103" 0.993m
#             magnets      = 10
#             Payout Scale factor= 3.910   0.099m
#             Tension Scale Factor= 200
#             operation limit= 5,980 lb
#
# Load pin in waterfall winches is sending out an A/D
# value of 2 even under 0 tension
# Also, payout pos/neg is opposite other winches
# uwf winch    sheave diam= 12.125   .308m
#             wire diam  = 0.322    .00818m
#             total circumference= 39.103" 0.993m
#             magnets      = 10
#             Payout Scale factor= 3.910   0.0993m
#             Tension Scale Factor= 60
#             operation limit= 5,980 lb
#
# lwf winch    sheave diam= 12.125   0.308m
#             wire diam  = 0.3125  0.00794m
#             total circumference= 39.074" 0.993m
#             magnets      = 10
#             Payout Scale factor= 3.907   0.0993m
#             Tension Scale Factor= 60
#             operation limit= 6,565 lb
#
# wnc1 and WNC2 are old.
#
# winch   payout tension speed
# name fields are in format A:B where y=Ax+B
#
# new winch strings
# meters out = mout * a

```

```

# speed = speed * c
# tension = (tension * b) - e
#      a      b      c      d      e
SWNC -0.1   200   1.67  20718  -800
PWNC  0.1   180   1.67  20150   0
BWNC  0.1   62.5  1.67   5980  437.5
WWNC -0.1   60   -1.67  5980   0
# old winch strings
LWF -0.1   60   -1.67  6565   0
UWF -0.1   60   -1.67  5980   0
WNC1  0.1   200  1.67  NAN   0
WNC2 -0.1   60  -1.67  NAN   0
#
#
#####
#####
##
***** Calibration factors for SBE 21 S/N 1390 *****
***** Calibration Date of 12 Mar 05 *****
# c:currently in use
# Temperature calibration factors
%TEMPERTURE%
g 0.00421050756
h 0.00059535523
i 0.00000497876949
j -0.00000173798388
fo 1000.000
*

# conductivity calibration factors
%CONDUCTIVITY%
g -3.93226726
h 0.470569719
i 0.000634631789
j -0.00000987772523
p -0.000000095700
t 0.0000032500
*

***** Remote Temperature Probe SN #1497 *****
***** Calibration Date of 10-Nov-05 *****
# external temperature calibration factors
%EXTERNAL TEMPERATURE%
g 0.00473766449
h 0.000668793556
i 0.0000284645709
j 0.00000261326034
fo 1000.000
*

#
#####

```